



## Instructions for Use

**REF** Great White Platinum Stirrups 800-0342-PL

### *Replacement Pads*

Platinum Stirrup Boot Pads, Set 508-1415

### **INTENDED USE**

Great White Platinum Stirrups are a patient positioning accessory used in gynecology, urology, and laparoscopic procedures. They provide a safe system for positioning the legs of up to 600 lb. (272 kg) patients in a variety of lithotomy/abduction positions. The intended users of this device are medical professionals within hospitals and surgery centers.

### **GENERAL SPECIFICATIONS**

#### *Device Dimensions (maximum)*

- Height: 40.25" +/- 0.5" (102 cm +/- 1 cm)
- Width: 9" +/- 0.5" (23 cm +/- 1 cm)
- Depth: 14" +/- 0.5" (36 cm +/- 1 cm)
- Device Weight Per Stirrup: 11 +/- 0.5 lbs. (5 +/- .22 kg) (with pad)
- Range of Motion: -30° to 75° lithotomy range, -9° to 25° adduction to abduction
- Single-person installation
- Connects to seat section of surgical table
- Store stirrups on transport/storage dolly (P/N 800-0074-S sold separately)

### **GENERAL INFORMATION**

- Product not made with Natural Rubber Latex
- Device supports 600 lbs. (272 kg) proportional patient load (6'4" (193 cm) tall patient per 99% human body model)
- Product warranty covers product from manufacturing defects for period of 2 years
- If damaged in shipping, call Customer Service at (888) 724-8763 or (781) 982-7000 to obtain Return Material Authorization number. For product warranty issues, contact Customer Service.
- CE marked medical device according to MDR (EU) 2017/745
- Life of device is 5 years under normal use
- Store device between -4°F to +86°F (-20°C to 30°C)

### **INSTRUCTIONS**

Become familiar with patient positioning device's features before use with patient. Always practice on a nurse, physician or appropriate volunteer prior to using clinically.



1. Attach SpringLoc Clamps P/N 800-0338 on accessory rails in same location on opposite sides of surgical table at patient's hip joints.

2. Prior to placing device into SpringLoc Clamp, identify patient's left and right side stirrup indicated on boot label.



3. Insert stirrup blades into clamps. Tighten clamps by turning handle to the right.



4. Ensure patient is positioned on surgical table in accordance with procedure and surgeon requirements.

**WARNING!!** To prevent patient or operator injury from inadvertent stirrup movement, securely tighten accessory clamp and boot clamp.



(1) Trigger Handle (2) Free Floating Boot Pivot (3) Boot Clamp

5. Support patient's leg by grasping heel in one hand and underside of knee with other hand. Gently flex knees and transfer leg into boot, then secure boot straps.

6. To achieve appropriate leg/foot position, loosen boot clamp. Adjust boot to desired position and re-tighten clamp. Ensure patient's heels are securely seated in boot heels.

7. To achieve appropriate lithotomy and abduction positions, squeeze trigger, adjust to desired position and release to lock.

**Note:** Free-floating boot is designed to rotate about pivot shown above through-out the full range of motion, reducing risk of superficial nerve injury.

## TAKE DOWN

1. Loosen clamps and remove stirrups by lifting them out of clamps
2. Remove accessory clamps from side rails

## COMPONENT OVERVIEW

Great White Platinum Stirrups provide a safe system for positioning the legs of up to 600 lb. (272 kg) patients in a variety of lithotomy/abduction positions.

### ***Replacement Pads***

**508-1415** Platinum Stirrup Boot Pads, Set

*Other required products for use:*

**800-0338** SpringLoc Clamp

**800-0338-EU** SpringLoc Clamp

**800-0338-JPN** SpringLoc Clamp

**800-0338-SWISS** SpringLoc Clamp

## DISPOSAL

- **General** - Prevent infection by cleaning and disinfecting product before disposal
- **Packaging** - Dispose packaging material via household waste according to national requirements
- SchureMed accepts back used or retired products - or dispose of product in accordance with national requirements



## PRODUCT USE WARNINGS

**WARNING!** Maximum load should not exceed appropriate proportion of a patient weighing 600 lbs. (272 kg). Use care with low-maximum load capacity surgical tables that accessory rails are not overloaded.



**WARNING!** Do not reuse device if there are obvious signs of damage or functional issues. Consult manufacturer before reusing.



**WARNING!** Hazards result from incorrect use. Strictly follow instructions for use with your Operating Table system.



**WARNING!** Surgical table load capacities may be less. Never overload a surgical table. Device is intended for mounting on side of a rail of a surgical table only.

## CLEANING RECOMMENDATION

Follow current Association of periOperative Registered Nurses Journal Guidelines for proper cleaning and disinfection procedure.

**WARNING!**

*Adhere to standards for blood-borne pathogens from the Occupational Safety and Health Administration. Use recommended protective clothing, gloves, masks and eye protection to clean accessory.*

**CAUTION**

Strictly read/follow manufacturer's directions for cleaning fluids. DO NOT use cleaners containing phenolics.

- 1. Remove major contaminants from accessory with disposable materials. Follow appropriate bio-hazard waste disposal procedures.*
- 2. Apply cleaning fluid liberally to entire accessory and wipe with clean, lint-free cloth until all moisture and cleaning fluid is removed from accessory*
- 3. Let accessory dry*

**USER NOTICE**

*Any serious incident that has occurred in relation to the device should be reported to the manufacturer and the competent authority of the Member State in which the user and/or patient is established.*

**eIFU Language Versions**

To download and print the Instructions for Use, please go to [www.schuremed.com](http://www.schuremed.com)

**Symbol Glossary**

Symbol	Title
	Manufacturer
	Date of manufacture
	Authorized Representative in the European Community
	Authorized Representative in the Swiss
	Importer
	Serial Number
	Warning
	Medical Device
	Unique Device Identifier
	CE Marking



**SchureMed** (081001460)

452 Randolph Street, Abington, MA 02351 USA

**Toll Free** (888) 724-8763 | **Ph** (781) 982-7000 | **Fax** (781) 982-7001 | **orders@schuremed.com**

